

Certificate of Accuracy			
TPG57 & TPG60 Universal Service Gauge			
Issue: 1	Date: 27/09/2012	Author: SMN	Page: 1 of 1

The accuracy of each Universal Service Gauge when released from our factory is +/- 2 psi.

Each Tyre Pressure Gauge, before release, is checked and calibrated on test equipment that has accuracy traceable to the UK National Physical Laboratory reference No 0221 via our Druck DPI 150 Pressure Indicator s/n 2329290.



Steve Martin

QA Process Manager